



<https://www.revistaneurocirugia.com>

## O-FUN-28 - Vocal accelerometer recording of TMS evoked language disturbances: preliminary study for awake surgery application

J. Leote<sup>1,2,3,4</sup>, J. Valls-Sole<sup>4</sup>, J. Costa<sup>5,6</sup>, J. Casanova<sup>4</sup>, L. Oleaga<sup>7</sup>, J. Fonseca<sup>8</sup>, R.G. Nunes<sup>3</sup> and H.A. Ferreira<sup>3</sup>

<sup>1</sup>Neurosurgery Department, Hospital Garcia de Orta, Almada. <sup>2</sup>Faculty of Medicine, University of Lisbon. <sup>3</sup>Instituto de Biofísica e Engenharia Biomédica, Faculty of Health Sciences, University of Lisbon. <sup>4</sup>EMG Unit, Neurology Department, Hospital Clinic, Universitat de Barcelona, IDIBAPS, Barcelona. <sup>5</sup>Laboratory of Clinical Pharmacology and Therapeutics, Faculty of Medicine, University of Lisbon. <sup>6</sup>Center for Evidence-Based Medicine, Faculty of Medicine, University of Lisbon. <sup>7</sup>Neuroradiology Department, Hospital Clinic, University of Barcelona. <sup>8</sup>Language Research Laboratory, Institute of Molecular Medicine (IMM), Faculty of Medicine, University of Lisbon.

### Resumen

**Introduction:** The effect of repetitive navigated transcranial magnetic cortical stimulation (r-nTMS) on language as a method to prevent damage during surgery involving speech-relevant brain sites remains a crude, mostly qualitative, aspect of neuromonitoring.

**Objectives:** To evaluate if uniaxial accelerometer recording of vocal infraglottic vibration is a feasible method for quantitative assessment of the expected interference of r-nTMS on language.

**Material and methods:** T1-weighted magnetic resonance images were acquired in 8 healthy and were fed into a computer providing directional space projection of a figure of '8' magnetic coil used for r-nTMS. A uniaxial accelerometer was placed over the cricothyroid cartilage. We applied 5 pulses of 20Hz at 80% of resting motor threshold over the opercular inferior frontal gyrus (opIFG) of the left cerebellar hemisphere. Subjects performed a picture naming-task paradigm. In test trials, a train of r-nTMS was applied at the same time and 200 ms after the picture display. We analyzed off-line frequency, onset latency and amplitude of the accelerometer recording.

**Results:** The subjects were submitted to 50 test and 20 control pictures randomly presented. Vocal fundamental frequency was lower in males ( $121 \pm 14$  Hz) than in females ( $201 \pm 116$  Hz). Similar amplitudes and onset latencies were recorded. The r-nTMS yielded 11 clinical language disturbances (4 anomia and 6 paraphasias) mostly resulting from opIFG stimulation with 0ms delay of picture and n-rTMS.

**Conclusions:** Recording vibration of infraglottic structures with a uniaxial accelerometer is a useful method to assess language disturbances induced by brain stimulation. The method can be applied in the operating room to monitor surgery around language areas.